

Alpha Resources, Inc. Certificate Of Analysis

CYCLOHEXANONE-2,4-DINITROPHENYLHYDRAZONE

AEB-2005

Lot #: 12203

Carbon,Hydrogen & Nitrogen

Carbon = 51.79%
Hydrogen = 5.07%
Nitrogen = 20.14%


The values For CH&N were determined using a commercial CHN Analyzer against Acetanilide SRM 141c as offered by NIST.

Oxygen

Oxygen = 22.00%

The value for oxygen was determined using a commercial oxygen analyzer against Chlorobenzoic Acid available from NIST. This is SRM 2144.

Remedies for any claimed defect in this product will be limited to product replacement or refund of the purchase price. In no event shall Alpha Resources be liable for incidental or consequential damages.



Prepared by D.M. Reasons
Director of Quality Assurance