

Alpha Resources, Inc. Certificate Of Analysis

AR-1713

COAL STANDARD

LOT #LIG96

DRIED BASIS VALUE

WEIGHT PERCENT SULFUR = 1.18
STANDARD DEVIATION = 0.03

Methods
ASTM D 4239 C

NIST standards employed were SRM 2684 & 2685

Notes:

Normal ASTM procedures should be employed when using this standard. This includes using the *reproducibility* and *repeatability* factors. Use the reproducibility factor when establishing the first set of data, and then establish an internal repeatability factor as your guideline.

The statistical analysis and overall coordination leading to certification was performed by D.M. Reasons at Alpha Resources.

Remedies for any claimed defect in this product will be limited to product replacement or refund of the purchase price. In no event shall Alpha Resources be liable for incidental or consequential damages.

This standard was produced in accordance to Guide 34. Alpha Resources has become accredited under the ISO Guide 34:2009 for RMP and holds a ISO 17025 accreditation. Refer to certificate and scope of accreditation for details.

EXPIRATION DATE

THIS CRM IS VALID FOR TWO YEARS FROM THE DATE OF OPENING



P.O Box 199 3090 Johnson Road Stevensville, MI 49127 Phone (269) 465-5559 Fax (269) 465-3629

www.alpharesources.com

Page 1 Of 1