

Alpha Resources, Inc.

Certificate Of Analysis

AR-1720
PROXIMATE COAL STANDARD
LOT # 720309 LID # 775309
DRIED BASIS VALUE

MEAN WEIGHT % SULFUR	=	0.29
STANDARD DEVIATION	=	0.02
MEAN WEIGHT % ASH	=	6.23
STANDARD DEVIATION	=	0.2
MEAN WEIGHT % VOL. MATTER	=	43.25
STANDARD DEVIATION	=	0.8
BTU	=	11749
STANDARD DEVIATION	=	49
FIXED CARBON (calculated)	=	50.52

Methods Employed:

SULFUR	ASTM	D 4239
ASH	ASTM	D 3174, D5142
VOL. MATTER	ASTM	D 3175, D5142
BTU	ASTM	D 5865

Notes:


The material used in production of this standard was prepared in accordance with ARI 031. The precision values represent the standard deviation (k=1) obtained through analytical testing, and are given for reference only. Normal ASTM procedures should be employed when using this standard; this includes using the *reproducibility* and *repeatability* factors of the method for establishing analytical uncertainty. When necessary, professional judgment is applied toward consideration of data and statistical information. The statistical analysis and the overall direction and coordination of the analytical measurements leading to certification were performed by K.E. Dyer, Technical Manager, at Alpha Resources.

The samples for round robin testing were selected in accordance with ARI 014. The above values relate only to the material used to produce this standard. The analytical samples were dried under a nitrogen atmosphere for a minimum of 70 minutes at 107° C +/- 3°C until a steady mass is achieved. This reference standard was prepared in accordance to ISO Guide 34-2009. Store sealed under normal laboratory conditions. This bottle contains 50g, minus 60 mesh (250 micron) fine powder.

Remedies for any claimed defect in this product will be limited to product replacement or refund of the purchase price. In no event shall Alpha Resources be liable for incidental or consequential damages. This is a Certified Reference Material (working standard). For good laboratory practice it is recommended that all standards be verified prior to use.

EXPIRATION DATE
THIS CRM IS VALID FOR TWO YEARS FROM THE DATE OF OPENING

CERTIFIED April 10, 2012



P.O. Box 199 3090 Johnson Road Stevensville, MI 49127 USA Phone (269) 465-5559 Fax (269) 465-3629

www.alpharesources.com

Page 1 Of 1