

Alpha Resources, Inc.

Certificate Of Analysis

#2 DIESEL STANDARD

AR2075CP

LOT 075CP1193

ROUND ROBIN RESULTS

LABORATORY	CLOUD POINT ASTM D-2500 DEGREES F	POUR POINT ASTM D-97 DEGREES F
1	+ 8.5	- 3.0
2	+10.5	- 6.0
3	+14.5	- 5.0
4	+12.0	- 4.0
5	+ 8.0	- 3.0
	MEAN= +10.5	MEAN= - 4.0

Notes:

Before use, the contents of the bottle should be mixed through vigorous shaking. The contents should not be exposed to air for a lengthy period of time.

Normal ASTM procedures should be employed when using this standard. This includes using the reproducibility and repeatability factors. Use the reproducibility factor when establishing the first set of data, and then establish an internal repeatability factor as your guideline.

The statistical analysis of the certification was performed by D.M. Reasons at Alpha Resources.

The overall direction and coordination of the analytical measurements leading to certification were performed by D.M. Reasons at Alpha Resources.

Remedies for any claimed defect in this product will be limited to product replacement or refund of the purchase price. In no event shall Alpha Resources be liable for incidental or consequential damages.

EXPIRATION DATE

THIS CRM IS VALID FOR TWO YEARS FROM THE DATE OPENING



P.O Box 199 3090 Johnson Road Stevensville, MI 49127 Phone (269) 465-5559 Fax (269) 465-3629

www.alpharesources.com