

# Alpha Resources, Inc. Certificate Of Analysis

**AR-2781  
ULTIMATE COAL STANDARD**

**LOT# 706407**

**LID# 706407**

**DRIED BASIS VALUES**

**Proximate Analysis**

% Ash.....	17.75
% Volatile Matter.....	18.15
% Fixed Carbon.....	64.17
% Sulfur.....	1.92
Btu.....	12,030

**Ultimate Analysis**

% Carbon.....	70.33
% Hydrogen.....	3.40
% Nitrogen.....	1.35
% Chlorine.....	0.035
% Oxygen.....	5.45

**Sulfur Forms**

% Pyritic.....	0.71
% Organic.....	0.99
% Sulfate.....	0.12
% Total.....	1.82

**Ash Fusion Temperature**

Initial Deformation	Reducing / Oxidizing	(Degrees F)	2183 / 2514
Softening H = W	Reducing / Oxidizing		2240 / 2549
Softening H = ½ W	Reducing / Oxidizing		2330 / 2567
Fluid	Reducing / Oxidizing		2379 / 2595

**Mineral Analysis**

Silica.....	49.81
Alumina.....	21.84
Titania.....	1.03
Ferric Oxide.....	21.10
Calcium Oxide.....	0.77
Magnesium Oxide.....	0.78
Strontium Oxide.....	0.02

Potassium Oxide.....	2.46
Sodium Oxide.....	0.29
Sulfur Trioxide.....	0.68
Phosphorus Pentoxide.....	0.22
Manganese Oxide.....	0.01
Barium Oxide.....	0.10
Undetermined.....	0.50

The Following ASTM Procedures Were Employed In The Analysis Of All Samples:

Preparation.....	ASTM D 2013	Ash Analysis.....	ASTM D 3682
Carbon & Hydrogen.....	ASTM D 3178	Ash.....	ASTM D 3174
Volatile Matter.....	ASTM D 3175	Fusibilities.....	ASTM D 1857
Sulfur.....	ASTM D 3177	BTU.....	ASTM D 2015
Nitrogen.....	ASTM D 3179	Sulfur Forms.....	ASTM D 2492

The material used in production of this standard was sampled in accordance with ARI 031. The above values relate only to the material used to produce this standard.

The overall direction and coordination of the analytical measurements leading to certification were performed by J.R. Shingledecker at Alpha Resources.

Remedies for any claimed defect in this product will be limited to product replacement or refund of the purchase price. In no event shall Alpha Resources be liable for incidental or consequential damages.

This is a Certified Reference Material (CRM). For good laboratory practice it is recommended that all standards be verified prior to use.

**EXPIRATION DATE  
THIS CRM IS VALID FOR TWO YEARS FROM THE DATE OF PURCHASE WHICH  
APPEARS ON THE BOTTLE**

*J.R. Shingledecker*