

# Alpha Resources, Inc. Certificate Of Analysis

AR-2846

## KEROSENE STANDARD

LOT #461010

LID ID 461010

WEIGHT PERCENT SULFUR = 0.30  
STANDARD DEVIATION = 0.02

### High Purity Materials/Standards used:

T-Butyl Di-Sulfide

### Methods used for analysis:

ASTM D 5453

This standard was produced gravimetrically using high purity materials, with balances calibrated and checked by precision NIST traceable weights, and verified using the above ASTM test methods. The precision value represents the 95% confidence limit (k=2) derived from analysis and errors from weighing. When necessary, professional judgment is applied toward consideration of data and statistical information. The statistical analysis and the overall direction and coordination of the analytical measurements leading to certification were performed by K.E. Dyer at Alpha Resources. Normal ASTM procedures should be employed when using this standard. This includes using the *reproducibility* and *repeatability* factors for the ASTM method you wish to employ.

### Notes:

Before use, the contents of the bottle should be mixed through vigorous shaking. The contents should not be exposed to air for a lengthy period of time. This product has an indefinite shelf life until opened.

This is a Certified Reference Material (CRM), and is verified by the above-mentioned standards. For good laboratory practice it is recommended that all standards be verified prior to use. Remedies for any claimed defect in this product will be limited to product replacement or refund of the purchase price. In no event shall Alpha Resources be liable for incidental or consequential damages.

### EXPIRATION DATE

THIS CRM IS VALID FOR TWO YEARS FROM THE DATE OF OPENING

October 20, 2010

