Alpha Resources, Inc. Certificate Of Analysis

AR 302

CAST IRON STANDARD

LOT # 11402

% CARBON MEAN = 3.35 ONE SIGMA = 0.03 TWO SIGMA = 0.06 RANGE = 3.29 = 3.41

$ \text{TWO SIGMA} = 0.03 \\ \text{TWO SIGMA} = 0.06 \\ \text{RANGE} = 3.29 - 3.41 $	TWO SIGMA = 0.002 RANGE = 0.039 - 0.047
I) (I	Analysis is ASTM E 1019 Latest Revision) Standards Employed:
NIST IRSID CKD NCS	SRM 4i, 4j, 4L, 107c 502-1 227, 230 HC 16011, 56005 Notes:

The certified mean values were derived by a number of data sets provided by various laboratories over different days by separate analysts utilizing ASTM approved instruments.

The precision values represent the standard deviation, two times the standard deviation, and complete range of analyses.

The statistical analysis and the overall direction and coordination of the analytical measurements leading to certification were performed by K.E. Dyer at Alpha Resources.

Remedies for any claimed defect in this product will be limited to product replacement or refund of the purchase price. In no event shall Alpha Resources be liable for incidental or consequential damages.

This is a Certified Reference Material (CRM), and is traceable to the above-mentioned standards. For good laboratory practice it is recommended that all standards be verified prior to use.

Certified April, 2002

Kent Der

% SULFUR

MEAN = 0.043

P.O. Box 199 3090 Johnson Road Stevensville, MI 49127 USA Phone (269) 465-5559 Fax (269) 465-3629 www.alpharesources.com Page 1 Of 1