

Alpha Resources, Inc. Certificate Of Analysis

FLASH POINT STANDARD

#2 DIESEL OIL

AR 3025

LOT 0751099S288 Lid ID 07599

FLASH POINT = 162 DEGREES FARENHEIT
STANDARD DEVIATION = 5 DEGREES FARENHEIT

METHOD:ASTM D 93

Notes:

Before using the contents of this bottle should be mixed through vigorous shaking. The contents should not be exposed to air for a lengthy period of time.

Normal ASTM procedures should be employed when using this standard. This includes the reproducibility and repeatability factors. Use the reproducibility factor when establishing the first set of data, then establish an internal repeatability factor as your guide line.

The statistical analyses and the overall direction and coordination leading to certification was performed by D.M. Reasons at Alpha Resources.

Remedies for any claimed defect in this product will be limited to product replacement or refund of the purchase price. In no event shall Alpha Resources be liable for incidental or consequential damages.

EXPIRATION DATE

**THIS CRM IS VALID FOR TWO YEARS FROM DATE OF PURCHASE WHICH
APPEARS ON THE BOTTLE**



Alpha Resources, Inc.
Certificate Of Analysis

□

