

Alpha Resources, Inc.

Certificate Of Analysis

#1 DIESEL VISCOSITY STANDARD

AR 3028
LOT 028893

VISCOSITY (in cSt at 40 deg.C) = 2.43 0.08

ROUND ROBIN SUMMARY

LAB	VISCOSITY	METHOD
1	2.35	1
2	2.43	1
3	2.51	1
4	2.40	1
5	2.46	1

METHOD:

1. ASTM D-445

NOTES:

Before use, the contents of the bottle should be mixed through vigorous shaking. The contents should not be exposed to air for a lengthy period of time.

The statistical analysis of the certification was performed by D.M. Reasons at Alpha Resources.

The overall direction and coordination of the analytical measurements leading to certification were performed by D.M. Reasons at Alpha Resources.

Remedies for any claimed defect in this product will be limited to product replacement or refund of the purchase price. In no event shall Alpha Resources be liable for incidental or consequential damages.

This is a secondary standard reference material, and should be verified against a primary standard such as NIST when possible.

EXPIRATION DATE

THIS SRM IS VALID FOR TWO YEARS FROM DATE OF OPENING



P.O Box 199 3090 Johnson Road Stevensville, MI 49127 Phone (269) 465-5559 Fax (269) 465-3629

www.alpharesources.com

Page 1 Of 1