

Alpha Resources, Inc.

Certificate Of Analysis

AR 303
CAST IRON STANDARD
LOT # 21208

% CARBON
MEAN = 3.93
ONE SIGMA = 0.04
TWO SIGMA = 0.08
RANGE = 3.85 to 4.01

% SULFUR
MEAN = 0.0185
ONE SIGMA = 0.0008
TWO SIGMA = 0.0016
RANGE = 0.0172 to 0.0202

Method of Analysis is ASTM E 1019-03 and ARI 033
Primary Standards Employed:

NIST	107c, 338
NCS	HC11003, HC16011, NS56005
IPT	66
JSS	111-12

Notes:

The mean analytical values were derived by a number of data sets (n=50) by various instrumentation meeting ASTM E1019-03. The precision values represent the standard deviation, two times the standard deviation, and complete range of analysis. When necessary, professional judgment is applied toward consideration of data and statistical information. The statistical analysis and the overall direction and coordination of the analytical measurements leading to certification were performed by K.E. Dyer at Alpha Resources.


The material used in production of this standard was sampled in accordance with ARI 032. The samples for round robin testing were selected in accordance with ARI 014. The above values relate only to the material used to produce this standard.

Remedies for any claimed defect in this product will be limited to product replacement or refund of the purchase price. In no event shall Alpha Resources be liable for incidental or consequential damages.

This is a Certified Reference Material (CRM), and is traceable to the above-mentioned standards. For good laboratory practice it is recommended that all standards be verified prior to use.

Alpha Resources is an ISO/IEC 17025 accredited laboratory. For more information concerning our scope of accreditation contact Alpha Resources.

Certified June, 2008



P.O Box 199 3090 Johnson Road Stevensville, MI 49127 Phone (269) 465-5559 Fax (269) 465-3629

www.alpharesources.com

Page 1 Of 1