

Alpha Resources, Inc. Certificate Of Analysis

AR-3045

#2 DIESEL API

LOT # 451015

LID ID 751015

ASTM METHOD
D 4052

TEST
API @60F

RESULTS
35.1

Notes:

This bottle contains 500ml of #2 Diesel distillate to be used as per your test method. Kept sealed and unopened this product has an indefinite shelf life. Once opened this certificate is valid for two years. Keep sealed and stored under normal laboratory conditions, minimizing any exposure to air or light. The material used in production of this standard was sampled in accordance with ARI 041. The samples for round robin testing were selected in accordance with ARI 031. The above values relate only to the material used to produce this standard. This material was produced in accordance to ISO Guide 34-2009.

Remedies for any claimed defect in this product will be limited to product replacement or refund of the purchase price. In no event shall Alpha Resources be liable for incidental or consequential damages.

This is a Reference Material (RM), and is verified using the above-mentioned test method(s). For good laboratory practice it is recommended that all standards be verified prior to use.

EXPIRATION DATE

THIS RM IS VALID FOR TWO YEARS FROM THE DATE OF OPENING

Certified December 1, 2015



Kent Dyer, Technical Manager