

Alpha Resources, Inc. Certificate Of Analysis

HARDGROVE GRINDABILITY INDEX

ARHARDGROVE

LOT 10-98B

VALUES FOR BOTTLES CONTAINED IN SET

10-98-A	38
10-98-B	64
10-98-C	85
10-98-D	98

This set contains four standard coal samples for calibrating Hardgrove Grindability equipment. The Hardgrove Grindability Index for each sample is labeled on the container.

To calibrate the grindability machine, follow the procedure given in Standard Test Method D-409-93A, Annual Book of ASTM Standards.

Reprints of the standard test method are available from the American Society for Testing and Materials (A.S.T.M.) at the following address:

ASTM
1916 Race Street
Philadelphia, PA 19103

Remedies for any claimed defect in this product will be limited to product replacement or refund of the purchase price. In no event shall Alpha Resources be liable for incidental or consequential damages.

EXPIRATION DATE
THIS CRM IS VALID FOR ONE YEAR FROM THE DATE OF PURCHASE

