

Alpha Resources, Inc. Certificate Of Analysis

AR 1726

COAL STANDARD

LOT #726897 LID ID 170697

DRIED BASIS VALUE

MEAN WEIGHT % SULFUR	=	2.37
STANDARD DEVIATION	=	0.03
MEAN WEIGHT % ASH	=	11.59
STANDARD DEVIATION	=	0.08
MEAN WEIGHT % VOL. MATTER	=	35.41
STANDARD DEVIATION	=	0.81
BTU	=	11,550
STANDARD DEVIATION	=	50
MEAN WEIGHT % FIXED CARBON	=	52.95

Methods Employed:

SULFUR	ASTM	D 4239 METHOD C
ASH	ASTM	D 3174
VOL. MATTER	ASTM	D 3175
BTU	ASTM	D 3286 & 1989

Notes:

The certified mean values were statistically derived from the values reported to Alpha Resources through Round Robin.

Normal ASTM procedures should be employed when using this standard. This includes using the *reproducibility* and *repeatability* factors. Use the reproducibility factor when establishing the first set of data, and then establish an internal repeatability factor as your guideline.

The statistical analysis and overall coordination leading to certification was performed by D.M. Reasons at Alpha Resources.

Remedies for any claimed defect in this product will be limited to product replacement or refund of the purchase price. In no event shall Alpha Resources be liable for incidental or consequential damages.

EXPIRATION DATE

**THIS CRM IS VALID FOR TWO YEARS FROM THE DATE OF PURCHASE WHICH
APPEARS ON THE BOTTLE**



P.O. Box 199 3090 Johnson Road Stevensville, MI 49127 USA Phone (408) 700-0000 FAX (408) 700-0629

www.alpharesources.com

Page 1 Of 1