

Alpha Resources, Inc. Certificate Of Analysis

DPR 1044

COAL STANDARD

LOT #044201

DRIED BASIS VALUE

WEIGHT PERCENT SULFUR = 0.10
STANDARD DEVIATION = 0.01

The above value represent the mean values reported to Alpha Resources through Round Robin.

Methods

ASTM D 5373

ASTM D 4239

ASTM D 2683a

Notes:

Normal ASTM procedures should be employed when using this standard. This includes using the *reproducibility* and *repeatability* factors. Use the reproducibility factor when establishing the first set of data, and then establish an internal repeatability factor as your guideline.

The statistical analysis and overall coordination leading to certification was performed by D.M. Reasons at Alpha Resources.

Remedies for any claimed defect in this product will be limited to product replacement or refund of the purchase price. In no event shall Alpha Resources be liable for incidental or consequential damages.

EXPIRATION DATE

**THIS CRM IS VALID FOR TWO YEARS FROM THE DATE OF PURCHASE WHICH
APPEARS ON THE BOTTLE**



P.O. Box 199 3090 Johnson Road Stevensville, MI 49127 USA Phone (269) 465-5559 Fax (269) 465-3629

www.alpharesources.com

Page 1 Of 1