

Alpha Resources, Inc.

Certificate Of Analysis

KED1022
CATALYST STANDARD
LOT # 080405

% CARBON
MEAN = 25.5
ONE SIGMA = 0.3
TWO SIGMA = 0.6
RANGE = 24.9 – 26.1

% SULFUR
MEAN = 12.1
ONE SIGMA = 0.2
TWO SIGMA = 0.4
RANGE = 11.7 – 12.5

Primary Standards Employed:

High Purity Calcium Carbonate
High Purity Iron Sulfide
NCS 70020
IRSID 701-1

Notes:

The mean analytical values were derived by a number of data sets utilizing various ASTM approved instruments.

The precision values represent the standard deviation, two times the standard deviation, and complete range of analysis. When necessary, professional judgment is applied toward consideration of data and statistical information.

The statistical analysis and the overall direction and coordination of the analytical measurements leading to certification were performed by K.E. Dyer at Alpha Resources.

Remedies for any claimed defect in this product will be limited to product replacement or refund of the purchase price. In no event shall Alpha Resources be liable for incidental or consequential damages.

This is a Certified Reference Material (CRM), and is traceable to the above-mentioned standards. For good laboratory practice it is recommended that all standards be verified prior to use.

Certified August, 2005



Page 1 of 1