



Certificate of Analysis

AR-2864

RESIDUAL FUEL OIL CRM

LOT # 864220 LID ID 864220
% SULFUR = 29.04
EXPANDED UNCERTAINTY = ± 1.56
(k=2.1, @ 95% confidence limit, n=20)

Methods used for analysis:

ASTM D 1552-16, and ARI-035

Reference standards used: AR2863-861009, AR2862-862909, AR2864-864708, AR2863-863816

The intended use of this standard is for the calibration and or verification of sulfur analysis in petroleum-based materials by combustion IR detection or other valid testing methods. This standard was produced gravimetrically using high purity materials, with balances calibrated by precision NIST traceable weights, and verified using the above ASTM method. The uncertainty value represents the estimated expanded uncertainty using ANOVA, ISO Guide 35, and the Guide to Uncertainty Measurement. Metrological traceability is to the SI derived unit of mass fraction expressed as percent. The sample size used for analysis was approximately 50mg. Sample size and minimum sample size may be contingent upon your test method or instrumentation manufacturer recommendations. The material used in production of this standard was sampled in accordance with ARI 041. The samples for round robin testing were selected in accordance with ARI 031. When necessary, professional judgment is applied toward consideration of data and statistical information. The statistical analysis and the overall direction and coordination of the analytical measurements leading to certification were performed by K.E. Dyer at Alpha Resources. Normal ASTM procedures should be employed when using this standard. This includes using the reproducibility and repeatability factors for the ASTM method you wish to employ for overall method uncertainty. The above values relate only to the material used to produce this standard.

Before use, the contents of the bottle should be mixed by vigorous shaking. Any exposure to air and light should be kept to a minimum. Keep sealed and store upright under normal laboratory conditions. This bottle contains 100ml to be used as per the test method. While unable to determine a definite shelf life this reference should be reviewed 20 years from the date of certification. Once opened this certificate is valid for two years. This reference material was produced in accordance to ISO 17034. Remedies for any claimed defect in this product will be limited to product replacement or refund of the purchase price. In no event, shall Alpha Resources be liable for incidental or consequential damages. This certificate cannot be reproduced except in full.

This is a Certified Reference Material (CRM) and is traceable to the above-mentioned standards. For good laboratory practice, it is recommended that all standards be verified as fit for purpose prior to use. These test results are accredited under the Alpha Resources LLC laboratory's ISO/IEC 17025 and ISO 17034 (RMP) accreditation issued by ANSI-ASQ/ANAB. Refer to certificate and scope of accreditation(s) AT-1200 and AR1920.

EXPIRATION DATE

THIS CRM IS VALID FOR TWO YEARS FROM THE DATE OF OPENING

Certified February 28, 2020

Chief Chemist